


Welch Allyn FlexiPort[®] reusable cuff conversion instructions

BVL pack for wall/mobile aneroids or the **Hand pack** for hand aneroids

1. Remove all cuffs and attach tri-purpose connector to the aneroid or to the hand gauge.
2. Leave connector attached to the device and click cuffs on/off at the FlexiPort.

Monitor pack for automated blood pressure devices

1. Remove original hose connectors. Pull them out of the hose or cut them off with scissors.
2. Push in the one- or two-port FlexiPort connector respective to the number of tubes in the hose you are converting.
3. Leave connector attached to the device and click cuffs on/off at the FlexiPort.




Welch Allyn FlexiPort[®] reusable cuff conversion instructions

BVL pack for wall/mobile aneroids or the **Hand pack** for hand aneroids

1. Remove all cuffs and attach tri-purpose connector to the aneroid or to the hand gauge.
2. Leave connector attached to the device and click cuffs on/off at the FlexiPort.

Monitor pack for automated blood pressure devices

1. Remove original hose connectors. Pull them out of the hose or cut them off with scissors.
2. Push in the one- or two-port FlexiPort connector respective to the number of tubes in the hose you are converting.
3. Leave connector attached to the device and click cuffs on/off at the FlexiPort.




Welch Allyn FlexiPort[®] reusable cuff conversion instructions

BVL pack for wall/mobile aneroids or the **Hand pack** for hand aneroids

1. Remove all cuffs and attach tri-purpose connector to the aneroid or to the hand gauge.
2. Leave connector attached to the device and click cuffs on/off at the FlexiPort.

Monitor pack for automated blood pressure devices

1. Remove original hose connectors. Pull them out of the hose or cut them off with scissors.
2. Push in the one- or two-port FlexiPort connector respective to the number of tubes in the hose you are converting.
3. Leave connector attached to the device and click cuffs on/off at the FlexiPort.




Welch Allyn FlexiPort[®] reusable cuff conversion instructions

BVL pack for wall/mobile aneroids or the **Hand pack** for hand aneroids

1. Remove all cuffs and attach tri-purpose connector to the aneroid or to the hand gauge.
2. Leave connector attached to the device and click cuffs on/off at the FlexiPort.

Monitor pack for automated blood pressure devices

1. Remove original hose connectors. Pull them out of the hose or cut them off with scissors.
2. Push in the one- or two-port FlexiPort connector respective to the number of tubes in the hose you are converting.
3. Leave connector attached to the device and click cuffs on/off at the FlexiPort.



Welch Allyn FlexiPort reusable cuffs

- Markings on cuff help size cuff appropriately for each patient
- Anti-microbial coating
- 360° rotatable tubing helps you position your patient more easily
- Rated for at least 80,000 inflations
- 3-year warranty

Ordering information

| Number | Size | Range |
|----------|--------------|----------|
| Reuse-06 | Small Infant | 7-10 cm |
| Reuse-07 | Infant | 9-13 cm |
| Reuse-08 | Small Child | 12-16 cm |
| Reuse-09 | Child | 15-21 cm |
| Reuse-10 | Small Adult | 20-26 cm |

| Number | Size | Range |
|-----------|------------------|----------|
| Reuse-11 | Adult | 25-34 cm |
| Reuse-11L | Adult Long | 25-34 cm |
| Reuse-12 | Large Adult | 32-43 cm |
| Reuse-12L | Large Adult Long | 32-43 cm |
| Reuse-13 | Thigh | 40-55 cm |

Note: These cuffs come without tube sets and only work if you have Welch Allyn FlexiPorts attached to your devices/gauges.

To learn more about Welch Allyn FlexiPort Reusable cuffs, call Welch Allyn customer service at 1-800-535-6663 or visit Welch Allyn's website at www.welchallyn.com.



Welch Allyn, Inc.
4341 State Street Road
P.O. Box 220
Skaneateles Falls, NY USA 13153-0220

Mat'l No. 714837 Ver. A

Welch Allyn FlexiPort reusable cuffs

- Markings on cuff help size cuff appropriately for each patient
- Anti-microbial coating
- 360° rotatable tubing helps you position your patient more easily
- Rated for at least 80,000 inflations
- 3-year warranty

Ordering information

| Number | Size | Range |
|----------|--------------|----------|
| Reuse-06 | Small Infant | 7-10 cm |
| Reuse-07 | Infant | 9-13 cm |
| Reuse-08 | Small Child | 12-16 cm |
| Reuse-09 | Child | 15-21 cm |
| Reuse-10 | Small Adult | 20-26 cm |

| Number | Size | Range |
|-----------|------------------|----------|
| Reuse-11 | Adult | 25-34 cm |
| Reuse-11L | Adult Long | 25-34 cm |
| Reuse-12 | Large Adult | 32-43 cm |
| Reuse-12L | Large Adult Long | 32-43 cm |
| Reuse-13 | Thigh | 40-55 cm |

Note: These cuffs come without tube sets and only work if you have Welch Allyn FlexiPorts attached to your devices/gauges.

To learn more about Welch Allyn FlexiPort Reusable cuffs, call Welch Allyn customer service at 1-800-535-6663 or visit Welch Allyn's website at www.welchallyn.com.



Welch Allyn, Inc.
4341 State Street Road
P.O. Box 220
Skaneateles Falls, NY USA 13153-0220

Mat'l No. 714837 Ver. A

Welch Allyn FlexiPort reusable cuffs

- Markings on cuff help size cuff appropriately for each patient
- Anti-microbial coating
- 360° rotatable tubing helps you position your patient more easily
- Rated for at least 80,000 inflations
- 3-year warranty

Ordering information

| Number | Size | Range |
|----------|--------------|----------|
| Reuse-06 | Small Infant | 7-10 cm |
| Reuse-07 | Infant | 9-13 cm |
| Reuse-08 | Small Child | 12-16 cm |
| Reuse-09 | Child | 15-21 cm |
| Reuse-10 | Small Adult | 20-26 cm |

| Number | Size | Range |
|-----------|------------------|----------|
| Reuse-11 | Adult | 25-34 cm |
| Reuse-11L | Adult Long | 25-34 cm |
| Reuse-12 | Large Adult | 32-43 cm |
| Reuse-12L | Large Adult Long | 32-43 cm |
| Reuse-13 | Thigh | 40-55 cm |

Note: These cuffs come without tube sets and only work if you have Welch Allyn FlexiPorts attached to your devices/gauges.

To learn more about Welch Allyn FlexiPort Reusable cuffs, call Welch Allyn customer service at 1-800-535-6663 or visit Welch Allyn's website at www.welchallyn.com.



Welch Allyn, Inc.
4341 State Street Road
P.O. Box 220
Skaneateles Falls, NY USA 13153-0220

Mat'l No. 714837 Ver. A

Welch Allyn FlexiPort reusable cuffs

- Markings on cuff help size cuff appropriately for each patient
- Anti-microbial coating
- 360° rotatable tubing helps you position your patient more easily
- Rated for at least 80,000 inflations
- 3-year warranty

Ordering information

| Number | Size | Range |
|----------|--------------|----------|
| Reuse-06 | Small Infant | 7-10 cm |
| Reuse-07 | Infant | 9-13 cm |
| Reuse-08 | Small Child | 12-16 cm |
| Reuse-09 | Child | 15-21 cm |
| Reuse-10 | Small Adult | 20-26 cm |

| Number | Size | Range |
|-----------|------------------|----------|
| Reuse-11 | Adult | 25-34 cm |
| Reuse-11L | Adult Long | 25-34 cm |
| Reuse-12 | Large Adult | 32-43 cm |
| Reuse-12L | Large Adult Long | 32-43 cm |
| Reuse-13 | Thigh | 40-55 cm |

Note: These cuffs come without tube sets and only work if you have Welch Allyn FlexiPorts attached to your devices/gauges.

To learn more about Welch Allyn FlexiPort Reusable cuffs, call Welch Allyn customer service at 1-800-535-6663 or visit Welch Allyn's website at www.welchallyn.com.



Welch Allyn, Inc.
4341 State Street Road
P.O. Box 220
Skaneateles Falls, NY USA 13153-0220

Mat'l No. 714837 Ver. A