



Vitrex Examination Gloves

Powder-Free

1. Product Description

Vitex examination gloves are designed for high resistance, strong performance, good dexterity and promotes easy donning.

2. Product specification

- Manufactured from nitrile butadiene and Polyvinyl Chloride
- Manufactured in accordance with:
- European PPE Directive 89/686/EEC
- Tested to:
 - EN420
 - EN455 Part 1, 2, 3
 - EN374 Part 1
 - AQL 1.5
 - Food Safe
- Extractable Protein: <7.5⁺
- Shelf life of 3 years

3. Product benefits

- · Latex and powder-free reducing the risk of hypersensitivity and allergic reactions
- Low in residual chemicals to reduce the risk of contact dermatitis associated with these chemicals
- Flexible material for improved feel and fit reducing fatigue during extended wear
- Tear resistant providing greater durability compared to latex or vinyl
- Beaded cuff for added strength and easier donning
- Low protein

4. Dimensions

	Small	Medium	large
Palm Width (mm)	85 ± 5	95 ± 5	105 ± 5
Length (mm)	250 <u>+</u> 10	250 <u>+</u> 10	250 <u>+</u> 10
Thickness (mm) palm	0.11 ± 0.03	0.11 ± 0.03	0.11 ± 0.03
Thickness (mm) finger	0.10 ± 0.03	0.10 ± 0.03	0.10 ± 0.03
Thickness (mm) cuff	0.10 ± 0.03	0.10 ± 0.03	0.10 ± 0.03

EN420: glove measurements are determined by considering the behaviors of the glove material and intended use.

DATA SHEET



Vitrex Examination Gloves

Powder-Free

5. Packing					
inner: 100 gloves per dispenser outer: 10 dispensers per case					
6. Symbols					
Manufacturer	LOT Lot Number	Do not re-use	Latex-free		
Temperature Limitation	Expiry date	Keep dry	Recyclable packaging		
CE marking					